

A multicentric verification of hereditary and acquired thrombophilia coagulation assays

ABSTRACT

Aims: With an ever-growing number of patients being admitted to our University Hospital Sveti Duh (Zagreb, Croatia), there was a need to positively respond to clinicians' demands about hereditary and acquired thrombophilia testing.

Study design: Prior to implementation of new assays into everyday laboratory routine, an extensive and multicentric verification of coagulation assays included in thrombophilia testing was performed.

Place and Duration of Study: Department of Medicine Medical Laboratory Diagnostics, University Hospital Sveti Duh, Zagreb, Croatia, between January 2020 and December 2021.

Methodology: Verification protocol included the precision study, following the CLSI EP15-A3 protocol; trueness estimation by comparison with two large hospital laboratories with established thrombophilia testing, as well as verification of reference intervals and cut-off values, as defined in the CLSI EP28-A3C and in the CLSI HA-60 guideline.

Results: All of the obtained imprecision CVs were within the manufacturer's claims. Seventy remnant plasma samples were included in the inter-laboratory comparison of PC and AT activities. While the observed bias for PC was within the EFLM analytical performance specifications for bias, the average bias for AT was higher than the EFLM acceptance criteria. P&B regression also revealed a significant positive proportional difference, thus revealing the need for method recalibration. Comparison of 30 samples for PS activity yielded a high negative bias (-33.4%) that exceeded the acceptance criteria of 8.1%, which was confirmed with a significant negative proportional error. Regarding LA testing, we preferably decided to compare the final classification of all received samples. The diagnostic accuracy was 99% when compared to Sestremitosrdnice UHC (N=70) and 95% when compared to University Hospital Centre Zagreb (N=19). All manufacturer's reference intervals and cut-offs were verified.

Conclusion: The verification study confirmed all manufacturer's claims, except for PS which has been replaced with free PS:Ag determinations. Therefore, the assays subject to verification could be safely introduced into routine practice.

Keywords: thrombophilia, hereditary, acquired, verification, coagulation

1. INTRODUCTION

Thrombophilia testing has long been limited to specialized centres. However, due to the high annual incidence rate of venous thromboembolism of 1 to 2 cases per 1000 individuals, as well as automation of the majority of assays, there is an ever-increasing demand for thrombophilia screening. It has been documented that ready availability of thrombophilia tests, particularly for heritable disorders, leads to their excessive and inappropriate use [1]. However, with expansion and

an ever-growing number of patients being admitted to the Hematology and Obstetrics and gynecology departments in our University Hospital Sveti Duh (Zagreb, Croatia), there was a need to positively respond to clinicians' demands about hereditary and acquired thrombophilia testing. It was particularly challenging to implement the non-standardized coagulation assays, never before assayed in our coagulation laboratory. To assure high-end quality, up-to-date assays, a literature search was performed [2-6]. After the selection of assays, and prior to implementation of new assays into everyday laboratory routine, an extensive and multicentric verification of coagulation assays included in both hereditary and acquired thrombophilia testing was performed.

2. MATERIAL AND METHODS

2.1 Materials and methods

The following assays were verified on the BCS XP coagulation analyzer (Siemens Healthineers, Marburg, Germany): antithrombin activity (AT) (Innovance Antithrombin), protein C activity (PC) (Berichrom Protein C), protein S activity (PS) (Protein S Ac), free protein S antigen (free PS:Ag) (Innovance Free PS Ag), activated protein C resistance (APCR) (ProC Ac R and ProC Global + Coagulation Factor V Deficient Plasma), lupus anticoagulant (LA) screening (LA1 and activated partial thromboplastin time by using Dade Actin FSL as reagent) and confirmation test (LA2), factor VIII activity (FVIII) (Dade Actin FS and coagulation FVIII Deficient Plasma). All reagents were manufactured by Siemens Healthcare Diagnostics (Marburg, Germany). Verification protocol included the determination of within-run, between-run and total precision by analyzing Siemens Control plasma samples in the normal and pathological range (listed in Table 1.) for five consecutive days in quintuplicate, following the Clinical and Laboratory Institute (CLSI) EP15-A3 protocol [7], trueness estimation by comparison with two large hospital laboratories (Sestremitosrdnice University Hospital Center and University Hospital Centre Zagreb, Croatia) with established thrombophilia testing, as well as verification of reference intervals and cut-off values, as defined in the CLSI EP28-A3C and in the CLSI HA-60 guideline[8,9]. The Sestremitosrdnice University Hospital Center, as well as University Hospital Centre Zagreb utilized the same reagents as our Department. Remnant plasma samples from patients for whom thrombophilia screening was ordered were separated into clean plastic tubes without additives and no additional blood draw was performed for the purposes of this study. Samples were frozen at -35 °C and transported to our Department within two weeks from blood draw because of limited sample stability for LA samples. All samples were assayed in our institution immediately after delivery. Frozen plasma samples were thawed in a water bath at 37 °C, well stirred and assayed as a batch. Imprecision coefficients of variation (CVs, expressed in percentages) were compared to manufacturer's claims.

2.1 Statistical analysis

The comparison of methods was performed by Bland-Altman plot and Passing-Bablok regression. Biases, obtained from Bland-Altman analysis and expressed as percentages, were compared with minimum criteria available within the European Federation of Clinical Chemistry and Laboratory Medicine (EFLM) Biological Variation Database[10]. To assure independent evaluation of the assays under verification, during the verification period we have participated in the external quality control assessment scheme by the External quality Control of diagnostic Assays and Tests with a focus on Thrombosis and Haemostasis (ECAT) foundation. All data were stored in Microsoft Excel 2010 software. Statistical analysis was performed in MedCalc® v22.016 statistical software (MedCalc Software Ltd, Ostend; Belgium).

3. RESULTS AND DISCUSSION

3.1 Imprecision

Regarding imprecision study, all of the obtained imprecision CVs were within the manufacturer's claims, as shown in Table 1.

Table 1. Results of the total precision verification of a multicentric verification of hereditary and acquired thrombophilia coagulation assays

Assay (reagent)	(unit)	Coefficients of variations (CVs) / %			
		Control Plasma N (lot No. 507767A,	Control Plasma P (lot No. 556714A,	Manufacturer's claim	Within manufacturer's claim (YES/NO)

	exp. 20/05/2021)	exp. 06/01/2021)		
Antithrombin (%) (INNOVANCE® Antithrombin) (lot No. 00325, exp. 06/07/2022)	3.0	7.8	<10.0	YES
Protein C (%) (Berichrom® Protein C) (lot No. 49884, exp. 25/07/2021)	2.2	2.0	<10.0	YES
Protein S (%) (Protein S Ac) (lot No. 50009, exp. 21/10/2021)	8.1	8.4	<10.0	YES
Free PS:Ag (%) (INNOVANCE® Free PS Ag) (lot No. 00477, exp. 16/06/2023)	2.5	2.3	<10.0	YES
FVIII (%) (Dade® Actin FS and Coagulation Factor FVIII Deficient Plasma) (lot No. 547693, exp. 04/09/2021)	7.1	9.0	<15.0	YES
	Control Plasma N (lot No. 507767A, exp. 20/05/2021)	ProC Control (lot No. 524479A, exp. 14/10/2023)	Manufacturer's claim	
APCR (ProC® Ac R) (lot No. 49840, exp. 24/10/2020)	6.5	4.0	<10.0	YES
APCR (ProC® Global + Coagulation Factor V Deficient Plasma) (lot No. 00468, exp. 16/11/2022)	4.8	3.4	<5.0	YES
	LA Control Low	LA Control High	Manufacturer's claim	YES

	(lot No. 546097A, exp. 05/06/2021)	(lot No. 545967A, exp. 02/06/2021)		
LA1 Screening Reagent (s) (lot No. 567211A, exp. 31/12/2020)	2.6	3.2	<5.0	YES
LA2 Confirmation Reagent (s) (lot No. 548881A, exp. 11/01/2021)	2.2	2.1	<5.0	YES
	Control Plasma N (lot No. 507767A, exp. 20/05/2021)	Citrol 2 (lot No. 548502, exp. 19/04/2023)	Manufacturer's claim	
Dade® Actin FSL Activated PTT Reagent (lot No. 556996. exp. 11/01/2021)	4.7	2.5	<5.0	YES
APCR – activated protein C resistance				

3.2 Accuracy (method comparison study)

Seventy remnant plasma samples were included in the inter-laboratory comparison of PC and AT activities, spanning the whole linearity range (PC: 49 – 149%, AT: 34 – 150%) (Table 2.). While the observed bias for PC was within the EFLM analytical performance specifications for minimum allowed bias, the average bias for AT was higher than the EFLM acceptance criteria. Passing and Bablok regression also revealed a significant positive proportional difference (Table 2).

Table 2. Results of method comparison study of a multicentric verification of thrombophilia coagulation assays

Assay (unit) (reagent)	N	Comparison range	Bland-Altman analysis			Passing and Bablok regression			
			Average bias% (95% CI)	EFLM APS %	Within APS (YES/NO)	Intercept (95% CI)	Significant YES/NO	Slope	Significant YES/NO
Antithrombin (%) (INNOVANCE® Antithrombin)	70	34 – 150	+10.8 (8.5 – 13.0)	3.2	NO	-2.28 (-13.8 – 5.25)	NO	1.14 (1.05 – 1.27)	YES
Protein C (%) (Berichrom® Protein C)	70	49 – 149	+1.9 (0.9 – 2.8)	6.7	YES	-1.02 (-4.90 – 3.11)	NO	1.02 (0.98 – 1.07)	NO

Protein S (%) (Protein S Ac)	30	32 – 130	-33.4 (-29.0 – (-37.8))	8.1	NO	-0.37 (-26.65 – 18.86)	NO	0.62 (0.44 – 0.87)	YES
95% CI – Confidence Interval EFLM APS – analytical performance specifications for minimum allowed bias, according to EFLM Biological Variation Database									

The first received ECAT report revealed unsatisfactory result for AT activity with a positive bias (+9.8%) observed in comparison to other participants that used the same method, even though the biases of the commercial quality controls during the whole verification period were within the allowed manufacturer's claims ($\pm 20\%$). However, the received results were in accordance with the positive bias observed in the comparison study. With the second round of ECAT samples we recalibrated our assay, and re-analyzed the samples from the first round. The results were satisfactory, and we have received a certificate for AT activity with a negligible bias. Comparison of 30 samples for PS activity (range 32 – 130%) yielded a high negative bias (-33.4%) that exceeded the acceptance criteria of 8.1 %, which was also confirmed with a significant negative proportional error from Passing and Bablok regression (Table 2). During the verification period an instability of the reagent used was observed, with a need for recalibration with every batch. The on-board stability of the reagent is declared to be 2 hours, which put a lot of pressure on timely measurement of thawed samples. As the assay for free PS:Ag determination showed to be stable and reproducible, and the current recommendations [4] suggested that the free PS:Ag assay should be used as the initial assay of choice in the clinical coagulation laboratory, we have decided to dismiss the PS activity assay, and to implement the free PS:Ag assay, only.

Regarding LA testing, the two diagnostic centres from the comparison study report LA results in a different manner. Sestre milosrdnice University Hospital Center reports the non-standardized ratio as the final result of the screen-mix-confirm protocol, whereas University Hospital Centre Zagreb reports results dichotomously as positive/negative. We preferably decided to compare the final classification of all received samples, then to assess the comparability of the individual measurements of the screen-mix-confirm protocol. The diagnostic accuracy was 99% when compared to Sestre milosrdnice UHC (N=70) and 95% when compared to University Hospital Centre Zagreb (N=19) (Table 3).

Table 3. Diagnostic accuracy for lupus anticoagulant comparison of a multicentric verification of thrombophilia coagulation assays

N=19		Zagreb University Hospital Centre		
University Hospital Sveti Duh		LA positive	LA negative	Total
LA positive		9	0	9
LA negative		1	9	10
Total		10	9	19
Diagnostic accuracy = $\frac{TP + TN}{TP + FP + TN + FN} = \frac{18}{19} = 0.95 \times 100\% = 95\%$				
N = 70		Sestre milosrdnice University Hospital Center		
University Hospital Sveti Duh		LA positive	LA negative	Total
LA positive		4	0	4
LA negative		1	65	66
Total		5	65	70
Diagnostic accuracy = $\frac{TP + TN}{TP + FP + TN + FN} = \frac{69}{70} = 0.99 \times 100\% = 99\%$				

3.3 Verification of reference intervals

Finally, we performed a verification of reference intervals and diagnostic cut-offs in plasma samples of 20 healthy volunteers without any coagulation abnormality (male/female ratio = 8/12). The health status was checked with complete blood count (CBC) and C-reactive protein (CRP) measurements. Coagulation samples were drawn into BD Vacutainer® 9NC 0.105M 3.2% Buffered Sodium Citrate Glass 4.5 mL Blood Collection Tubes (BD, Plymouth, UK). All manufacturer's reference intervals and cut-offs were verified.

4. CONCLUSION

The verification study confirmed all manufacturer's claims, except for PS which has been replaced with free PS:Ag determinations. Therefore, the assays subject to verification could be safely introduced into routine practice. Indeed, starting in December 2021, we have successfully started with coagulation testing of hereditary and acquired thrombophilia in our Department.

CONSENT

Not applicable.

ETHICAL APPROVAL (WHEREEVER APPLICABLE)

Not applicable.

REFERENCES

1. Graham N, Rashid H, Hunt BJ. Testing for thrombophilia: clinical update. *British Journal of General Practice* 2014; 64 (619): e120-e122
2. Van Cott EM, Orlando C, Moore GW, Cooper PC, Meijer P, Marlar R. Recommendations for clinical laboratory testing for antithrombin deficiency; Communication from the SSC of the ISTH. *J Thromb Haemost* 2020;18:17-22
3. Cooper PC, Pavlova A, Moore GW, Hickey KP, Marlar RA. Recommendations for clinical laboratory testing for protein C deficiency, for the subcommittee on plasma coagulation inhibitors of the ISTH. *J Thromb Haemost* 2020;18:271-277
4. Marlar RA, Gausman JN, Tsuda H, Rollins-Raval MA, Brinkman HJM. Recommendations for clinical laboratory testing for protein S deficiency: Communication from the SSC committee plasma coagulation inhibitors of the ISTH. *J Thromb Haemost* 2020;19:68-74
5. Moore GW, Van Cott EM, Cutler JA, Mitchell MJ, Adcock DM. Recommendations for clinical laboratory testing of activated protein C resistance; communication from the SSC of the ISTH. *J Thromb Haemost* 2019;17:1555-1561
6. Devreese KMJ, de Groot PG, de Laat B, Erkan D, Favaloro EJ, Mackie I, et al. Guidance from the Scientific and Standardization Committee for lupus anticoagulant/antiphospholipid antibodies of the International Society on Thrombosis and Haemostasis. *J Thromb Haemost* 2020;18:2828-2839
7. CLSI. User Verification of Precision and Estimation of Bias; Approved Guideline – Third Edition. CLSI document EP15-A3. Wayne, PA: Clinical and Laboratory Standards Institute, 2014.
8. CLSI. Defining, Establishing, and Verifying Reference Intervals in the Clinical Laboratory; Approved Guideline—Third Edition. CLSI document EP28-A3c. Wayne, PA: Clinical and Laboratory Standards Institute; 2008.
9. CLSI. Laboratory Testing for the Lupus Anticoagulant; Approved Guideline. CLSI document H60-A. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.
10. Aarsand AK, Fernandez-Calle P, Webster C, Coskun A, Gonzales-Lao E, Diaz-Garzon J, Jonker N, Simon M, Braga F, Perich C, Boned B, Marques-Garcia F, Carobene A, Aslan B, Sezer E, Bartlett WA, Sandberg S. The EFLM Biological Variation Database. <https://biologicalvariation.eu/>, Accessed 6.11.2023.

