

Original Research Article

MODEL EYE PICTURES CAN LAY THE FOUNDATION FOR PICTURES OBSERVED BY PATIENTS HAVING CATARACT SURGERY WHILE SIMULATION

ABSTRACT:

Aim: The photographs of a model eye during the fake waterfall are to be evaluated by medical treatment. **Methodology:** This study was performed as a study in a research institution and a case procedure was carried out. Our current research was conducted at People Medical College Hospital, Nawabshah from April 2020 to March 2021. In the "front office" of a model eye, the iced rear surface was inserted with a flawed fogged focus, an acceptable intraocular focal point or an I/A tip. During the replicated Waterfall treatment, video pictures were collected from the back surface of the model eye. Before the waterfall surgery, 26 patients saw the video pictures and post-operatively rated their visual impressions. **Results:** View from behind were pictures of the movable focal parts, the I/A tip and the IOL add-on. The image via the opiated focus point and the IOL has been the light of the meticulous enlargement lens from behind without moving objects. However, after IOL added, pictures of the tiny and working space were switched off from behind by the light of the magnifying equipment. 70% of patients reported the visual feel of shifting focal components as video cuts and 58%. The overwhelming majority of patients advised that the film before their intended medical operation should be viewed. **Conclusion:** The model eye can re-enable visual insight for patients throughout the medical treatment for a waterfall. The preoperative view of the video pictures might aid to loosen the sufferers during the treatment.

Keywords: Model-Eye Pictures, Foundation Patients, Cataract Surgery, Simulation.

INTRODUCTION

More than 85% of individuals with a waterfall or vitreous medicine revealed visual sensation even during operation. The visual impressions also include evocative growth of white or shaded light on objects and devices, and fingers or hands of doctors. 65% of those who report seeing or seeing improvements. There were also protests which seemed like careful tools [1]. Though a significant number of descriptions were orally reported, one patient spoke about her waterfall cautiously through two creative projects, one consisting of pink and blue circles. The inventor hypothesized, with an imaginable stimulation of the shadowing photoreceptors by the ultrasonographic force, that these pictures indicate the development of a phacoemulsification test in the eye [2]. Nevertheless, individuals with extracapsular extractions conducted without ultrasound were reported as similar experiences. Another patient was an accomplished Craftsman, a PC designer who had had an intra-vitreous epiretinal membrane treatment. He compared his designs to the intraocular devices used in various phases of vitrectomy in general precisely [3]. Despite these accounts, due of the complexity of reconstruction of visuals, the exact instrument, the instrument that caused these sensations, was not explained. The greatest evidence to date is that the shadows created by tools in the previous room during waterfall therapy or vitrectomy are responsible for these symptoms [4]. Furthermore, it was revealed that the entopic pictures of the eye were derived from the shadow created by the tissue parts. However, throughout the distinct times of the waterfall procedure the precise photographs applied to the retina have not been clarified. The researchers then tried to find pictures that move

the retina and provide a contribution to the patients' visual insights during the therapy with Waterfall. To this end, we utilized a model eye with an ice rear surface and created framed pictures [5].

METHODOLOGY

Based on Gullstrand's natural eye model a model eye has been designed (Figure 1). The eye body was formed of 25 mm long metal hub. The focus point (cornea) had a power of 45.1 diopters with a center thickness of 0.6 mm. Our current research was conducted at People Medical College Hospital, Nawabshah from April 2020 to March 2021. The cornea's circular divergence was β 0,23 mm, which was the same as the normal round eye deviation. The dimensions of the pupil were 6.1 mm and the depth of the front pupil 6 mm. Simple polymethacrylate was used on the corneas and back of the miniature eye. The inside surface of the back was continually sands trawled with a pneumatic tension of 1 atm to allow pictures to be viewed from behind on a superficial level. The model eye was filled with appropriate salt layout and removed all air bubbles at room temperature. In the case of a medical waterfall technique, light from a lens was adjusted from the optical axis of classical eye to center of eye of the model. The eye was modified to vertically position the optical pivot point. The model eye has been placed beneath a computer-controlled camera and the frosty Focus of the frosted surface is focused. The photos cast on the elegant surface were also captured using unrivalled methodology (specialist sitting at 12 o'clock on individual's head) by a computerized camera during a mock waterfall medicine procedure. In the front room the images of the items viewed by the light lens with a camcorder on the lens, focused on the things in the front-most room, have been captured constantly. The strength of the xenon light of the lens was 0.8, which is frequently employed for treatment with medical waterfalls. In order for the expert to observe the model eye from 11 o'clock to deal with the position, the gentle magnification instrument has been built to switch the pictures of items in the anterior chamber that the specialist could see through the cautious magnification device to a model eye.

Figure 1: Gullstrand's natural eye model





Figure 2: The front chamber of silicone-colored focusing points recorded by the loupe and pictures created on rear surface of classical eye.



RESULTS

Figure 2 shows pictures in the front chamber of silicone-colored focusing points recorded by the loupe and pictures created on rear surface of classical eye. The picture generated by the opaque silicones at the rear of the model eye was that of the lens' light. The picture of lens light was more diffuse with opaque silicone focus point. An unpolished 28-gage needle was pushed in a circular motion on the exterior of a silicone focus point as in circle capsulorhexis via an opening in the rhythm of a cornea. The video assessment of background pictures didn't display an image or developments of the needle. A loyal light picture was noticed in the silent photographs from behind by the independent white focus point (Figure 2). The cross-figured light from the perforations between the two focus points was recognized as the illumination light of the lens. The direction of the light going through the pitches between the separated focal points demonstrated that the pictures from behind represented perfectly what was viewed by the carefully lensed lens. The back picture was an image of the lens from the lens that was scanned from the foggy center by a sharper edge with the photographs (Figure 3). At the point where the lens was switched off, a picture of the lens and the roof of the lit workplace could be seen from the back camera. After the lens was removed and an expert peered into the model eye, a specialist image was visible, and the video pictures showed the model eye in the space light. These photos were captured in the right way and were not inverted outside the model eye.

Figure 3: The back picture of the lens that was scanned from the foggy center by a sharper edge with the photographs




Table 1: Questionnaire survey (Patients response on model eye).

Question 1. How identical was the simulator video at the start of the operation (clip 1) to the one you saw during the deletion of the cataract lens and when it said, "you may feel it

somewhere in the eye?	
1. Different	13 eyes (65%)
Similar	8 eyes (33%)
3. Do not know	1 eye (7%)
4. Same	2 eye (14%)
Question 2. Did you see any light or images during the surgery?	
1. No	10 eye (0%)
2. Yes	22 eyes (100%)
Question 3. Which way did you observe your eye entering into the shaft-like content?	
1. Nasal side	8 eyes (40%)
2. Do not know	2 eye (10%)
3. Temporal side	10 eyes (50%)

DISCUSSION

Our findings demonstrated that items in addition the tools in anterior chamber observed on frozen posterior surface as defocused blunt multitudes or weak waves lit with the light of the loudspeaker. The designs on the tips of such devices were not as crisp as during the glass process. In addition, the pictures in the model eye that were closer to the back were not as vivid as the imitation vitrectomy [6]. The increasing distance between the equipment in the previous chamber and the back surface has caused the pictures on the back surface to blur. We examine pictures (by the trousers of the split focus point) from rear surface of the model eye in phakic, pseudophakia and aphakic [7]. The rear pictures were comparable, showing that occurrence of IOL in the front chamber does not disturb the vision that the cases were taking through the lens during medical process [8]. The pictures of the lit workroom were more clearly apparent when the IOL was inserted and light from the lens was destroyed and gone from the vision center of the model eye. The splendid light of the expanding lens therefore most probably caused pictures outside of the eye to form. This also happened during the recreated medical operation intracameral focal components and devices [9]. The visual meetings of patients were characterized as a blend of close but out of sight pictures of things and also featured items and structures on corneal surface of cornea also the inner eye. Movement things, such as liquids and air bags on corneal and eye surfaces, eye moving tools, and even aspirated focal portions emulsified are probably responsible for altering tones and forms of certain patients [10].

CONCLUSION

In view of all this, the visual impressions of waterfall medicine may be recreated in the model eye. Even the items lodged in the anterior chamber during the waterfall test are able to alter the perception of pictures by the surrounding eye components. The photographic pictures projected onto the rear of the model's eye and individual's visual practice in medical process are comparable, in any case. Viewing video pictures before surgery can assist to minimize patients' anxiety of visual meetings during the treatment.

Ethical Approval:

As per international standard or university standard written ethical approval has been collected and preserved by the author(s).

Consent

As per international standard or university standard, patients' written consent has been collected and preserved by the author(s).

COMPETING INTERESTS

Authors have declared that no competing interests exist.

REFERENCES:

1. Abdolrahimzadeh, S., Parisi, F., Plateroti, A. M., Evangelista, F., Fenicia, V., Scuderi, G., et al. (2019). Visual acuity, and macular and peripapillary thickness in high myopia. *Curr. Eye Res.* 42, 1468–1473. doi: 10.1080/02713683.2017.134769
2. Anwar, S. M., Majid, M., Qayyum, A., Awais, M., Alnowami, M., and Khan, M. K. (2018). Medical image analysis using convolutional neural networks: a review. *J. Med. Syst.* 42:226
3. Barañano, A. E., Wu, J., Mazhar, K., Azen, S. P., and Varma, R. (2018). Visual acuity outcomes after cataract extraction in adult latinos. The Los Angeles latino eye study. *Ophthalmology* 115, 815–821. doi: 10.1016/j.optha.2007.05.052
4. Chang, L., Pan, C. W., Ohno-Matsui, K., Lin, X., Cheung, G. C., Gazzard, G., et al. (2019). Myopia-related fundus changes in Singapore adults with high myopia. *Am. J. Ophthalmol.* 155, 991–999.e1.
5. Chen, S. C., Chiu, H. W., Chen, C. C., Woung, L. C., and Lo, C. M. (2018). A novel machine learning algorithm to automatically predict visual outcomes in intravitreal ranibizumab-treated patients with diabetic macular edema. *J. Clin. Med.* 7:475. doi: 10.3390/jcm7120475
6. Chung, Y. W., Choi, M. Y., Kim, J. S., and Kwon, J. W. (2019). The association between macular thickness and axial length in myopic eyes. *Biomed. Res. Int.* 2019:8913582.
7. Fu, H., Cheng, J., Xu, Y., Zhang, C., Wong, D. W. K., Liu, J., et al. (2018). Disc-aware ensemble network for glaucoma screening from fundus image. *IEEE Trans. Med. Imaging* 37, 2493–2501. doi: 10.1109/tmi.2018.2837012
8. Gao, X., Lin, S., and Wong, T. Y. (2015). Automatic feature learning to grade nuclear cataracts based on deep learning. *IEEE Trans. Biomed. Eng.* 62, 2693–2701. doi: 10.1109/tbme.2015.2444389

9. Gohil, R., Sivaprasad, S., Han, L. T., Mathew, R., Kiouisis, G., and Yang, Y. (2019). Myopic foveoschisis: a clinical review. *Eye (Lond.)* 29, 593–601. doi: 10.1038/eye.2014.311
10. Grassmann, F., Mengelkamp, J., Brandl, C., Harsch, S., Zimmermann, M. E., Linkohr, B., et al. (2018). Deep learning algorithm for prediction of age-related eye disease study severity scale for age-related macular degeneration from color fundus photography. *Ophthalmology* 125, 1410–1420. doi: 10.1016/j.optha.2018.02.037

UNDER PEER REVIEW